


PULSE GENERATOR

PERFORMANCE CHECK

AVO-3A-P

1567

NOV 21 83



- a) Output signal amplitude:
25 to 75 AMPS.
- b) Pulse width:
2.5 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

20 AMPS/DIV

2.0 NSEC/DIV

$Z_L = 5 \Omega$

+15 ± 10 V

250 mA (MAX)