


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AP-AN-HV2-05-N-M*

S.N.: *1582*

Date: *OCT 31 83*



a) Output signal amplitude:

*0 TO -20V*

b) Pulse width:

*0 TO 2 nsec*

c) Rise time:

*< 100 psec*

d) Fall time:

*< 200 psec*

e) PRF:

*0 TO 1 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+15 ± 1.0V*

*250 mA (MAX)*

*10 VOLTS/DIV*

*1.0 nsec/DIV*

*400 kHz*