

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVP-19U-14V2-05-N-M

S.N.:

1584

Date:

OCT 31 83



a) Output signal amplitude:

0 TO -20 V

b) Pulse width:

0 TO 2 NSEC

c) Rise time:

< 100 PSEC

d) Fall time:

< 200 PSEC

e) PRF:

0 TO 1.0 MHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 V

250 mA (MAX)

10 VOLTS/DIV

1.0 NSEC/DIV

400 KHz

ALL