


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PRK-A-1-CA-PS-N*

S.N.: *1593*

Date: *JAN 23 84*



-) Output signal amplitude:
0 TO -200 V
-) Pulse width:
50 nsec TO 1.0 μsec
-) Rise time:
≤ 10 nsec
-) Fall time:
≤ 10 nsec
-) PRF:
0 TO 100 KHz
(0.5% MAX DUTY CYCLE)
-) Jitter, stability:
OK
-) Prime power:

50 VOLTS/DIV
100 NSEC/DIV

R_L = 50 Ω

4 KHz

120/240 V
50-60 Hz

JA

