


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-P*
S.N.: *1597*
Date: *JAN 10 84*

- a) Output signal amplitude:
+25 TO +100 AMPS
- b) Pulse width:
25 NSEC
- c) Rise time:
N/A
- d) Fall time:
N/A
- e) PRF:
0 TO 1 KITE
- f) Jitter, stability:
OK
- g) Prime power:



*= 64 AMP/DIV
10 NSEC/DIV
1 KITE.*

*+15 ± 1.0 V
200 mA (MAX)
*[Signature]**