


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVI-P

S.N.: 1608

Date: JAN 30/84



- a) Output signal amplitude:  
0 TO +40V
- b) Pulse width:  
2 TO 100 NS
- c) Rise time:  
≤ 500 PSEC
- d) Fall time:  
≤ 2 NS
- e) PRF:  
500 Pk 1.2
- f) Jitter, stability:  
OK
- g) Prime power:

2 NS/DIV  
32 VOLTS/DIV  
40 kHz

+15 ± 1.0V  
200 mA MAX