


PULSE GENERATOR

PERFORMANCE CHECK

Model: *APV P-AV-1-05-M-17*

S.N.: *1610*

Date: *OCT 31 83*



- a) Output signal amplitude:  
*0 TO +10 V*
- b) Pulse width:  
*0 TO 4 μSEC*
- c) Rise time:  
*< 100 pSEC*
- d) Fall time:  
*< 135 pSEC*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 V/DIV  
1.0 μSEC/DIV  
400 kHz*

*+15 ± 1.0 V  
250 mA (MAX)  
*[Signature]**