

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AVN-1-C-PN-M

S.N.:

1611

Date:

NOV 30 83



a) Output signal amplitude:

0 TO $\pm 5V$

b) Pulse width:

0.2 TO 1.0 nsec

c) Rise time:

≤ 100 psec

d) Fall time:

≤ 135 psec

e) PRF:

50 TO 250 MHz

f) Jitter, stability:

OK

g) Prime power:

5.0 VOLT/DIV.
1.0 nsec/DIV.

110/220 V
50-60 Hz.