


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-51-C-05-P

S.N.: 1615

Date: DEC 23 83



a) Output signal amplitude:

0 TO +10V

b) Pulse width:

≤ 130 PSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK

g) Prime power:

110 / 220V

10 VOLTS/DIV.

1.0 NSEC/DIV

500 KHz

50-60 Hz