


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-C-P*

S.N.: *1621*

Date: *NOV 18 83*



- a) Output signal amplitude:
0 TO \approx 10V
- b) Pulse width:
0 TO \approx 1 NS
- c) Rise time:
 \leq 100 PSEC
- d) Fall time:
 \leq 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS / DIV
1.0 NS / DIV
400 kHz*

*110/220 V
50-60 Hz*