

PULSE GENERATOR
PERFORMANCE CHECK

Model:


ANP-AN-1411-C-P

S.N.:

1623

Date:

JAN 13 1984



a) Output signal amplitude:

0 to +15 V

b) Pulse width:

0 to 2 nsec

c) Rise time:

≤ 100 psec

d) Fall time:

≤ 150 psec

e) PRF:

0 to 1 MHz

f) Jitter, stability:

OK

g) Prime power:

10 VOLT/DIV
100 nSEC/DIV
400 kHz

120 / 240 V
50-60 Hz