


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVMR-1-125-PW-PN-MID*

S.N.: *1671*

Date: *FEB 28 1984*



- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
10 TO 200 NS
- c) Rise time:
≤ 1 NS
- d) Fall time:
≤ 1 NS
- e) PRF:
0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
5 NS/DIV
1 MHz

110/220V
50-60 Hz
[Signature]