


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ACMR-1-PS-PN-05-MPT*

S.N.: *1673*

Date: *Feb 28 1984*



- a) Output signal amplitude:
0 to ±5 Volts
- b) Pulse width:
10 to 200 nsec
- c) Rise time:
≤ 1 nsec
- d) Fall time:
≤ 1 nsec
- e) PRF:
0 to 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*5 Volts/DIV
5 nsec/DIV*

1 MHz

*110/220V
50-60 Hz*

[Signature]