

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVI-V-P

S.N.:

1674

Date:

MAR 12 84



a) Output signal amplitude:

0 TO 450 V.

b) Pulse width:

0 TO 100 nSEC

c) Rise time:

≤ 300 pSEC

d) Fall time:

≤ 2 nSEC

e) PRF:

0 TO 20 kHz

f) Jitter, stability:

OK

g) Prime power:

415 ± 1.0 V

25 mA (max)

32 VOLTS/DIV

5 nSEC/DIV

20 kHz