


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-3-CA-ED-P*

S.N.: *1675*

Date: *FEB 27 84*



- a) Output signal amplitude:  
*0 TO +15V*
- b) Pulse width:  
*5 TO 100 NSEZ*
- c) Rise time:  
*≤ 100 PSEZ*
- d) Fall time:  
*≤ 200 PSEZ*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS/DIV  
5 NSEZ/DIV  
100 KHz*

*+ 24 ± 1.0V  
200 mA (mA)*

*[Signature]*