


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-AV1-C-P*

S.N.: *1678*

Date: *JUN 16 1984*



*10 VOLTS/DIV
1.0 NS/DIV
400 KHz.*

- a) Output signal amplitude:
0 TO +10V
- b) Pulse width:
0 TO 4 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240V
50-60 Hz*

[Signature]