


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PK-P*

S.N.: *1680*

Date: *NOV 30 83*



a) Output signal amplitude:

*7.5 V*

b) Pulse width:

*4 to 100 nsec*

c) Rise time: (see file 1)

*≤ 100 psec*

d) Fall time:

*≤ 2 nsec*

e) PRF:

*see file 1, 2*

f) Jitter, stability:

*OK*

g) Prime power:

*5 VOLTS/DIV.*

*100 nSEC/DIV*

*40 kHz*

*+15 ± 1.0 V*

*150 mA (MAX)*

*[Signature]*