


PULSE GENERATOR

PERFORMANCE CHECK

Model: PVK-17

S.N.: 1681

Date: NOV 30 83



- a) Output signal amplitude:
 $> +5V$
- b) Pulse width:
470 100 nsec
- c) Rise time: (SEE FILE 1)
 ≤ 100 psec
- d) Fall time:
 ≤ 2 nsec
- e) PRF:
SEE FILE 1, 2
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
1.0 nsec/DIV.
40 KHz.

$+15 \pm 1.0 V$
150 mA (MAX)
M