


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-1-P.

S.N.: 1683

Date: MAR. 5, 1984.



- a) Output signal amplitude:
25 to 100 AMPS
- b) Pulse width:
25 nSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

= 63 AMPS/DIV
10 nSEC/DIV
1 KHz

+15 ± 10V
200 mA (MAX)
AJ