


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AGM-1-C-D*
S.N.: *1688*
Date: *FEB 27 84*



- a) Output signal amplitude:
0 TO +5V
- b) Pulse width:
0 TO 6 nsec
- c) Rise time:
≤ 100 psec
- d) Fall time:
≤ 135 psec
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
110/220V
50-60 Hz

5 V 0.5 nsec/div
1.0 nsec/div
5 MHz