


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANM-2-V*

S.N.: *1703*

Date: *JAN 24 84*



- d) Output signal amplitude:
0 TO -15V
- e) Pulse width:
0 TO 2 NSEC
- f) Rise time:
≤ 100 PSEC
- g) Fall time:
≤ 135 PSEC
- h) PRF:
0 TO 25 MHz
- i) Jitter, stability:
OK
- j) Prime power:

*10 VOLTS/DIV
1.0 NSEC/DIV
4 MHz.*

*+24 ± 1.0 V
500 mA (max)*

[Signature]