


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMH-14V1-P*

S.N.: *1704*

Date: *MAR 21 54*



a) Output signal amplitude:

0 TO +100V

b) Pulse width:

≤ 1 nsec

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 100 kHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0V
250 mA (MAX)

100 VOLTS/DIV
2.0 nsec/DIV
40 kHz