


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AUP-194-1-PS-N-TTL-05

S.N.: 1705

Date: MAR 7 '84



- a) Output signal amplitude:  
0 TO - 10V
- b) Pulse width:  
0 TO 4 NS
- c) Rise time:  
≤ 100 PS
- d) Fall time:  
≤ 135 PS
- e) PRF:  
0 TO 1 MHz
- f) Jitter, stability:  
OK
- g) Prime power:

10 VOLTS/DIV  
1.0 NS/DIV  
400 KHz.

110/220 V  
50-60 Hz