

PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-S-1-P

S.N.: 1706

Date: JAN 30/84

a) Output signal amplitude:

0 TO 110V

b) Pulse width:

≤ 130 PSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK

g) Prime power:

$\pm 15 \pm 1.0$ V

200 mA

(MAX)

10 VOLTS/DIV

100 nSEC/DIV

400 kHz

