


PULSE GENERATOR

PERFORMANCE CHECK

Model: *DVP-AU-HV3-C-PN-CA-EW-OS-M*

S.N.: *1710*

Date: *APRIL 16 84*



- a) Output signal amplitude:
0 TO ±40 V
- b) Pulse width:
0 TO 2 nSEC
- c) Rise time:
≤ 150 pSEC
- d) Fall time:
≤ 250 pSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
110/220 V
50 - 60 Hz

32 VOLTS/DIV
1.0 nSEC/DIV
100 KHz