


PULSE GENERATOR
PERFORMANCE CHECK

Model: MK-V-N-TTL
S.N.: 1714
Date: MAR 21 84



- a) Output signal amplitude:
7.5 V
- b) Pulse width:
6 TO 50 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
5 nSEC/DIV
20 KHz

+15 ± 1.0V
250 mA (MAX)
[Signature]