


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PV0-5-C-P*

S.N.: *1717*

Date: *MAY 19 84*



- a) Output signal amplitude:
0 TO +28 AMPS
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
≤ 3 NSECS
- d) Fall time:
≤ 4 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

≈ 32 AMPS/DIV
($R_L ≈ 3 \Omega$)
10 NSEC/DIV
4 KHZ.

110/220V
50-60HZ

NY