


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVI-V-1113 P*

S.N.: *1718*

Date: *APRIL 2 84*



*100 VOLTS/DIV  
10 NS/CDIV  
5 KHz*



*100 VOLTS/DIV  
1.0 NS/CDIV  
5 KHz*

a) Output signal amplitude:

*0 TO +100 V.*

b) Pulse width:

*0 TO 100 NS/CDIV.*

c) Rise time:

*< 200 PSEC*

d) Fall time:

*< 3 NS/CDIV*

e) PRF:

*0 TO 5 KHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+15 ± 1.0 Volts*

*300 mA (MAX)*