


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-AV-HV2-P*

S.N.: *1719*

Date: *APRIL 2 1980*



- d) Output signal amplitude:
0 TO +20V.
- e) Pulse width:
0 TO 2 NSEC
- f) Rise time:
≤ 100 PSEC
- g) Fall time:
≤ 200 PSEC
- h) PRF:
0 TO 1 MHz
- i) Jitter, stability:
OK
- j) Prime power:

10 VOLTS/DIV
1.0 NSEC/DIV
400 KHz.

+15 ± 1.0 VOLT.
250 mA (MAX)