


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-MP-N-05*

S.N.: *1720*

Date: *MAR 12 84*



- a) Output signal amplitude:  
*0 TO - 40V*
- b) Pulse width:  
*1 TO 100 nSEC*
- c) Rise time:  
*≤ 300 pSEC*
- d) Fall time:  
*≤ 2 nSEC*
- e) PRF:  
*BLE F16 1+2*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*5 nSEC/DIV*

*32 VOLTS/DIV*

*40 KHz*

*+15 ± 1.0 V*

*250 mA (MAX)*