


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-V-NC*

S.N.: *1725*

Date: *March 27 84*



- a) Output signal amplitude:
0 to 50 V (50-5)
- b) Pulse width:
0 to 100 nsec
- c) Rise time:
≤ 300 psec
- d) Fall time:
≤ 2 nsec
- e) PRF:
0 to 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:

32 VOUT/DIV
5 nsec/DIV
20 kHz
V_{cc} = 50 volts

110/220V
50-60 Hz