


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-1-C-PN-D-M-EP-EW-ED*

S.N.: *1735*

Date: *MAR 28 84*



- a) Output signal amplitude:  
*0 TO ± 5 VOLTS.*
- b) Pulse width:  
*0 TO 6 ~~μ~~NSER*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*5 VOLTS/DIV  
1.0 NSEC/DIV  
5 MHz*

*110/220V  
50-60 Hz*