


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-3-C-PN-M*

S.N.: *1740*

Date: *APRIL 23 1984*



*10 VOLTS/DIV
2 NSEC/DIV
5 MHz*

- a) Output signal amplitude:
0 TO ±15 VOLTS
- b) Pulse width:
2 TO 15 NSEC
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 600 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*110/220V
50-60 Hz*

[Signature]