


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-AV-1S-C-PN

S.N.: 1752

Date: APR. 10, 1984



d) Output signal amplitude:

0 to $\pm 5V$

e) Pulse width:

0 to 4 nsec

f) Rise time:

≤ 60 psec

g) Fall time:

≤ 135 psec

h) PRF:

0 to 1000 Hz

i) Jitter, stability:

OK

j) Prime power:

110/220 V

50-60 Hz

5 VOLTS/DIV

1.0 nsec/DIV

100 KHz