


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-1-PS-EA-P*

S.N.: *1953*

Date: *max 20 84*



a) Output signal amplitude:

*0 TO 200V*

b) Pulse width:

*50 nsec TO 5 usec*

c) Rise time:

*≤ 10 nsec*

d) Fall time:

*≤ 10 nsec*

e) PRF:

*0 TO 1 KHz*

f) Jitter, stability:

*OK*

g) Prime power:

*110/220V*

*50-60 Hz.*

*50 VOLTS/DIV*

*200 nsec/DIV*

*1 KHz*