


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-A-C-P*

S.N.: *1766*

Date: *MAR 21 84*



- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
4 TO 100 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
SEE PAGES 1-2
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV
2 NSEC/DIV
100 KHz*

*110/220V
50-60 Hz*