


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-U-C-P*

S.N.: *1768*

Date: *MAR 21 84*



*5 VOLTS/DIV
1.0 NS/DIV
40 KHz*

- a) Output signal amplitude:
≥ +5V
- b) Pulse width:
0 TO 50 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*110/220V
50-60 Hz*

[Signature]