


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-DMR1

S.N.: 1782

Date: FEB 16 84



- a) Output signal amplitude:
0 TO - 100 V
- b) Pulse width:
2 TO 100 NSER
- c) Rise time:
≤ 1 NSER.
- d) Fall time:
≤ 2 NSER
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

100 VOLTS/DIV

5 NSER/DIV

(CABLE SUPPLIED
WITH UNIT)

(CUT TO LENGTH
TO YIELD DESIRED
PW)

+ 15 ± 1.0 V

200 mA (MAX)