


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVN-2-P*

S.N.: *1783*

Date: *APRIL 27 1984*



- a) Output signal amplitude:
0 TO +2V
- b) Pulse width:
0.2 TO 1.5 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 100 pSEC
- e) PRF:
50 TO 250 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+24 ± 1.0 V
400 mA (MAX)

3.2 VOLTS/DIV
1.0 nSEC/DIV
200 MHz