


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-5-1-N-A

S.N.: 1785

Date: MAY 30 84



- a) Output signal amplitude:
0 TO -10V
- b) Pulse width:
≤ 130 P. SEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 kHz
- f) Jitter, stability:
OK
- g) Prime power:

1.0 nsec/div
10 volts/div
400 kHz

+15 ± 1.0 VOLT.
200 mA (max)

A handwritten signature or set of initials, possibly 'JH' or similar, written in dark ink.