


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-2-C-D

S.N.: 1790

Date: DMAR 19 84



- a) Output signal amplitude:  
0 TO + 350V
- b) Pulse width:  
0 TO 100 NSEC
- c) Rise time:  
 $\leq 1.5$  NSEC
- d) Fall time:  
 $\leq 3$  NSEC
- e) PRF:  
0 TO 5 KHZ
- f) Jitter, stability:  
OK
- g) Prime power:

10 NSEC/DIV  
320 VOLTS/DIV  
4 KHZ

110/220V  
50-60 HZ