


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVN-1-C-P*

S.N.: *1799*

Date: *APRIL 30 1984*



- a) Output signal amplitude:  
*0 TO +5V*
- b) Pulse width:  
*0.2 TO 1.0 NSEC.*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 100 PSEC.*
- e) PRF:  
*50 TO 250 MHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*110/220V*  
*50-60 Hz*

*3.2 VOLTS/DIV*  
*1.0 NSEC/DIV*  
*≈ 200 MHz (PRF)*

*[Signature]*