


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVN-1-C-N*

S.N.: *1800*

Date: *APRIL 30 1984*



-) Output signal amplitude:
0 TO -5V
-) Pulse width:
0.2 TO 1.0 NSEC
-) Rise time:
≤ 100 PSEC
-) Fall time:
≤ 100 PSEC
- e) PRF:
50 TO 250 MHz
- f) Jitter, stability:
OK
- g) Prime power:

3.2 VOLTS/DIV
1.0 NSEC/DIV
≈ 200 MHz (PRF)

110/220V
50-60 Hz

[Signature]