


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-MV-HV3-C-PN-M.*

S.N.: *1210*

Date: *APRIL 16 84*



- a) Output signal amplitude:  
*0 TO ± 40V*
- b) Pulse width:  
*0 TO 2 NSCZ.*
- c) Rise time:  
*≤ 150 PSCZ*
- d) Fall time:  
*≤ 250 PSCZ*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:
- g) Prime power: *OK*

*32 V/DIV*  
*1.0 NSCZ/DIV*  
*100 KHz.*

*110/220V*  
*50-60 Hz*