


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-V-14V3-C-PN-M*

S.N.: *1811*

Date: *APRIL 23 1984*



a) Output signal amplitude:

*0 TO ± 100 V*

b) Pulse width:

*0 TO 100 NS*

c) Rise time:

*≤ 200 PS*

d) Fall time:

*≤ 3 NS*

e) PRF:

*0 TO 5 KHz*

f) Jitter, stability:

*OK*

g) Prime power:

*110/220 V*

*50-60 Hz*

*100 VOLT/DIV*

*5 NS/DIV*

*5 KHz*

*JAC*