


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-3 -PN*
S.N.: *1814*
Date: *MAY 21 1984*



- a) Output signal amplitude:
0 TO ± 7 VOLTS
- b) Pulse width:
2 TO 15 NSEC
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 600 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
2 NSEC/DIV
4 MHz

+24 TO +30V
500 mA (MAX)

[Signature]