


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVI-U-141-C-DN-05-M-N*

S.N.: *1833*

Date: *MAY 8 84*



*10 NSEC/DIV
32 VOLTS/DIV
20 KHZ*



*1.0 NSEC/DIV
32 VOLTS/DIV
20 KHZ*

- a) Output signal amplitude:
0 TO ± 50 VOLTS
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
≤ 250 PSEC
- d) Fall time:
< 2 NSEC
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
*110/220V
50-60 HZ*

AG