


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-P*

S.N.: *1854*

Date: *APRIL 2 1984*



- d) Output signal amplitude:
+10 to +50 AMPS
- e) Pulse width:
≤ 25 nsec
- f) Rise time:
N/A
- g) Fall time:
N/A
- h) PRF:
0 to 1 kHz
- i) Jitter, stability:
OK
- j) Prime power:

*25 AMPS / DIV
10 nsec / DIV
1 kHz*

+15 ± 1.0 V

200 mA (max)