


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-2-N*

S.N.: *1855*

Date: *MAY 28 84*



- a) Output signal amplitude:  
*0 TO -15V...*
- b) Pulse width:  
*0 TO 2 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 2.5 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS/DIV      +24 ± 1.0 V*  
*1.0 NSEC/DIV      500 mA (MAX)*  
*4 MHz*