


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-2-N*
S.N.: *1865*
Date: *MAY 28 84*



- a) Output signal amplitude:
0 TO -125 V.
- b) Pulse width:
5 2 NSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 V

100 VOLT/DIV
2.0 NSEC/DIV
4 KHz

+15 ± 1.0 V
150mA (MAX)
[Signature]